## **INFRARED SYSTEMS DEVELOPMENT**CORPORATION



## ACQUIRE BODY TEMPERATURE WITH INFRARED CAMERA IN-SCENE REFERENCE

- **Φ CONTROLLED TO HUMAN BODY TEMPERATURE**
- **₱** REFERENCE SOURCE IN-SCENE WITH SUBJECT
- **Φ COMPENSATION OF ENVIRONMENTAL EFFECTS**
- Φ REMOVE THERMAL CAMERA ERROR
- Φ CALIBRATION OF THERMAL CAMERAS
- **DESCRIPTION HUMIDITY AND AMBIENT SENSORS**

Infrared Systems Development has teamed with Santa Barbara Infrared to offer this new low cost solution for real-time body temperature measurement systems. The Nightingale Reference Target source is designed for in-scene comparison of human body temperature. Using the Nightingale in the same scene as the subject allows for immediate visual and reporting verification of the Infrared Camera Temperature. Infrared Cameras and spot sensors can be affected by many factors in the environment which can be minimized with the Nightingale.

## NIGHTINGALE SPECIFICATIONS AND DIMENSIONS

## Specifications:

**Temperature Range: 30 - 45 °C (86-113°F)** 

Wavelength Range: 3 - 14 um

Emissivity: >0.95

Emitter Size: in (mm): 3" x 3" (76.2 x 76.2 mm)

Source Type: Extended Area

Temperature Uniformity: +/- 0.15 °C (+/- 0.3 °F) Center 1.5 x 1.5" Area

Calibration Accuracy: +/- 0.15 °C (+/- 0.3 °F) Radiometric

Stability: +/- 0.05 °C (+/- 0.1 °F)

Response Time: < 5 Minutes

Set Temperature Resolution: 0.1 °C (0.2 °F)

Control Type: Active Multi-Band P.I.D.

AC Adapter Line Voltage: 90 to 125 or 208-240 VAC 50-60 Hz

Power Requirements: 18 VDC, 1A AC Power Adapter Provided

Operating Environment: 22° C +3/- 5°C (71.6°F +5.4/- 9°F)

5%-90% R.H. N.C.

Storage Temperature Range: -20°C to 70°C (-4°F to 158°F)


Warranty: 1 Year


Remote Interface: USB


Mounting: 1/4 - 20 Tripod Mount

Weight: 1.5 lbs (.682 Kg)

Model Number: BTR-03







Specifications subject to change without prior notice

INFRARED SYSTEMS DEVELOPMENT CORPORATION

7319 Sandscove Court #4
Winter Park, FL 32792
Phone: (407) 679-5101 Fax: (407) 679-5520
sales@infraredsystems.com
www.infraredsystems.com